



Certificate of Compliance

Certificate: 1125320

Master Contract: 205078 (205078)

Project: 70045068

Date Issued: 2016-06-27

Issued to: Analog Devices Inc.
804 Woburn St
Wilmington, Massachusetts 01887-3462
USA

Attention: Mark Cantrell

The products listed below are eligible to bear the CSA Mark shown with adjacent indicator ▲



Issued by: *Martin Buchanan*
Martin Buchanan, P. Eng.

PRODUCTS

CLASS - C907330 - ELECTRONIC COMPONENTS-Optoisolators

Component Acceptance of Optoisolator Like Magnetic Coupling Devices:



Certificate: 1125320
Project: 70045068

Master Contract: 205078
Date Issued: 2016-06-27

Device (SOIC 8N R package)	Ratings		Standard/Notice and Clauses	Internal		External
	kV	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
ADuM1100xRZ ADuM3100xRZ	2.5	125	CSA CA5A, 14-13 Tbl 35, 6.2.1.1, 6.2.1.1/6.2.10, 6.8.1.1 60950-1-07+A1+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 61010-1-12+A1 K.3, K.4, 6.7.1.3, 6.7.2.2.2 A.17, K.6x1.6, K.7x1.6, 10 IEC 60950-1 2nd Ed.,+A1+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 61010-1 3 rd Ed A1 K.3, K.4, 6.7.1.3, 6.7.2.2.2 A.17, K.6x1.6, K.7x1.6, 10 EN 60950-1:2006+A1+A12+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2	-	-	4.0

where x may be A, B or U differentiating devices with different performance grades. Z, if used, indicates the lead-free version. Suffix EP at the end of the base part number indicates lower extended temperature capability to -55C with a NiPdAu lead finish. Further suffix letters or digits differentiate shipping package formats.

Notes:

1. These devices meet basic insulation requirements for 400Vrms for CSA 60950-1-07+A1+A2 and IEC 60950-1 2nd Ed.+A1+A2.
2. These devices meet reinforced insulation requirements for 148Vrms for CSA 60950-1-07+A1+A2 and IEC 60950-1 2nd Ed.+A1+A2. (pollution degree 2, material group III)
3. For CSA 61010-1-12+A1 and IEC 61010-1 3rd Ed. for 300Vrms, overvoltage category III for basic insulation and 150Vrms, overvoltage category IV for basic insulation, the devices meet clauses K.3, K.4, 6.7.1.3, 6.7.2.2.2 A.17, K.6x1.6, K.7x1.6 and 10. The risk management process is not applicable to these clauses. (pollution degree 2, material group III)
4. Case material: Multiple mold compounds have been evaluated. All meet at least material group II.
5. Evaluated by thermal cycling and other tests.
6. The creepage and clearance has been evaluated for altitudes ≤ 2000m, in pollution degree 2 and overvoltage category II except where specified above. (pollution degree 2, material group III).

These devices are Component Accepted as components for use in other Certified equipment where the suitability of the combination shall be determined by investigation in the final application.

APPLICABLE REQUIREMENTS

Component Acceptance Notice 5A (CA 5A) - Announcement of Extension of the Component Acceptance Service for Optocouplers and Related Devices
 CAN/CSA-C22.2 No 14-13 - Industrial Control Equipment



Certificate: 1125320
Project: 70045068

Master Contract: 205078
Date Issued: 2016-06-27

CAN/CSA-C22.2 No 60950-1-07+A1+A2 - Information Technology Equipment - Safety - Part 1:
General Requirements (Bi-national Standard, with UL 60950-1)
IEC 60950-1 2nd Ed.+A1+A2 - Information Technology Equipment - Safety - Part 1: General Requirements
EN 60950-1:2006+A1+A12+A2 - Information Technology Equipment - Safety - Part 1: General Requirements

Clauses K.3, K.4, 6.7.1.3, 6.7.2.2.2 A.17, K.6x1.6, K.7x1.6, 10
of

CAN/CSA-C22.2 No. 61010-1-12+A1 2015 - Safety Requirements for Electrical Equipment for
Measurement, Control and Laboratory Use - Part 1: General Requirements
(Tri-national standard, with UL 61010-1 Ed. 3 and ANSI/ISA-61010-1
(82.02.01)

and

IEC 61010-1:2010 3rd Ed. - Safety Requirements for Electrical Equipment for Measurement, Control
and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 1125320

Master Contract: 205078 (205078)

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70045068	2016-06-27	Update to 1125320 for alternative mould compound and leadframe based on results for similar devices from other report(s)
1936092	2007-08-20	Increase of working voltage to 400V
1709602	2005-09-16	Addition of Model ADuM3100XX-YYY
1438541	2003-05-27	To cover alternative outer case compound EME-6600H