



Certificate of Compliance

Certificate: 70112135

Master Contract: 255632

Project: 70112135

Date Issued: 2017-10-25

Issued to: **Linear Technology Corporation**
1630 McCarthy Blvd
Milpitas, California 95035
USA
Attention: Keith Bennett

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C', 'US' and ▲



Issued by: *Martin Buchanan*
Martin Buchanan, P. Eng.

PRODUCTS

CLASS - C907330 - ELECTRONIC COMPONENTS-Optoisolators

CLASS - C907391 - ELECTRONIC COMPONENTS-Optoisolators - Certified to US Standards

Component Acceptance of Isolator Module Devices:

Device (BGA 24 package)	Ratings		Clauses of Standard/Notice	Internal		External
	kV	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
LTM2894wx	7.5	125	CSA 60950-1-07+A1+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 62368-1-14 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.1.5.3, 4.5.8, 5.4.9.1, 4.1.2 60601-1:14 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7 IEC 60950-1 2nd Ed.,+A1+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 62368-1-14 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.1.5.3, 4.5.8, 5.4.9.1, 4.1.2 60601-1 Ed.3+A1 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7	-	-	18.1

Certificate: 70112135

Project: 70112135

Master Contract: 255632

Date Issued: 2017-10-25

Device (BGA 24 package)	Ratings		Clauses of Standard/Notice	Internal		External
	kV	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
			UL UL 1577 Standard for Optical Isolators			

where w is a letter for temperature range up to 125C, x is V for LGA package or Y for BGA package.

Notes:

1. These devices meet basic insulation requirements for 1810Vrms for CSA 60950-1-07+A1+A2 and IEC 60950-1 2nd Ed.+A1+A2. (pollution degree 2, material group III)
2. These devices meet reinforced insulation requirements for 905Vrms for CSA 60950-1-07+A1+A2 and IEC 60950-1 2nd Ed.+A1+A2. (pollution degree 2, material group III)
3. These devices meet basic insulation requirements for 1810Vrms for CSA 62368-1-14 and IEC 62368-1:2014. (pollution degree 2, material group III)
4. These devices meet reinforced insulation requirements for 905Vrms for CSA 62368-1-14 and IEC 62368-1:2014. (pollution degree 2, material group III)
5. For CSA 60601-1:14 and IEC60601-1 Ed.3+A1 for 2MOPP for 567.2Vrms, the devices meet clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7. The risk management process is not applicable to these clauses.
6. Case material: CTI=600V, erosion depth 0.028mm, meets material group I.
7. Evaluated by thermal cycling and other tests for a temperature rating of 125C.
8. The creepage and clearance has been evaluated for altitudes ≤ 2000m, in pollution degree 2, material group III and overvoltage category II except where specified otherwise.

These devices are Component Accepted as components for use in other Certified equipment where the suitability of the combination shall be determined by investigation in the final application.

APPLICABLE REQUIREMENTS

- | | | |
|--|---|---|
| CAN/CSA-C22.2 No 60950-1-07+A1+A2 | - | Information Technology Equipment - Safety - Part 1: General Requirements (Bi-national Standard, with UL 60950-1) |
| CAN/CSA-C22.2 No. 62368-1-14 | - | Audio/video, information and communication technology equipment– Part 1: Safety requirements |
| IEC 60950-1 2 nd Ed.+A1+A2 | - | Information Technology Equipment - Safety - Part 1: General Requirements |
| IEC 62368-1:2014 Ed. 2 | - | Audio/video, information and communication technology equipment– Part 1: Safety requirements |
| Clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7 | | |
| of | | |
| CAN/CSA C22.2 60601-1:14 | - | Medical Electrical Equipment Part 1: General requirements for basic safety and general performance (Adopted IEC60601-1:2005 Edition 3.0 +Amendment 1, 2012-07, MOD) |



Certificate: 70112135

Project: 70112135

Master Contract: 255632

Date Issued: 2017-10-25

and

IEC60601-1:2005 Ed 3.0+A1

- Medical Electrical Equipment Part 1: General requirements for basic safety and general performance

ANSI/UL 1577-2015

- Standard for Optical Isolators



Supplement to Certificate of Compliance

Certificate: 70112135

Master Contract: 255632

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70112135	2017-10-25	Original Component Acceptance