



Certificate of Compliance

Certificate: 80008667

Master Contract: 255632

Project: 80008667

Date Issued: 2019-11-29

Issued To: Analog Devices Incorporated
804 Woburn St.
Wilmington, Massachusetts, 01887-3462
United States

Attention: Keith Bennett

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C', 'US' and▲



Issued by: Martin Buchanan
Martin Buchanan, P. Eng.

PRODUCTS

CLASS - C907330 - ELECTRONIC COMPONENTS Optoisolators and non-optocal isolating devices

CLASS - C907391 - ELECTRONIC COMPONENTS Optoisolators and non-optical isolating devices -
Certified to US Standards

Component Acceptance of Isolator Module Devices:

Device (BGA LGA 36 package)	Ratings		Clauses of Standard/Notice	Internal		External
	kV	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
LTM2810wx-z	7.5	125	CSA 60950-1-07+A1+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 62368-1-14 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.1.5.3, 4.5.8, 5.4.9.1, 4.1.2 60601-1:14 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7 IEC	-	-	16.0

Certificate: 80008667

Project: 80008667

Master Contract: 255632

Date Issued: 2019-11-29

Device (BGA LGA 36 package)	Ratings		Clauses of Standard/Notice	Internal		External
	kV	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
			60950-1 2nd Ed.,+A1+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 62368-1-14 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.1.5.3, 4.5.8, 5.4.9.1, 4.1.2 60601-1 Ed.3+A1 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7 UL UL 1577-2015 Standard for Optical Isolators - Single protection			

where w is a letter for temperature range up to 125C, x is V for LGA package or Y for BGA package and z is blank or I for I²C or S for SPI data option.

Notes:

1. These devices meet basic insulation requirements for 1600Vrms for CSA 60950-1-07+A1+A2 and IEC 60950-1 2nd Ed.+A1+A2. (pollution degree 2, material group III)
2. These devices meet reinforced insulation requirements for 800Vrms for CSA 60950-1-07+A1+A2 and IEC 60950-1 2nd Ed.+A1+A2. (pollution degree 2, material group III)
3. These devices meet basic insulation requirements for 1600Vrms for CSA 62368-1-14 and IEC 62368-1:2014. (pollution degree 2, material group III)
4. These devices meet reinforced insulation requirements for 800Vrms for CSA 62368-1-14 and IEC 62368-1:2014. (pollution degree 2, material group III)
5. For CSA 60601-1:14 and IEC60601-1 Ed.3+A1 for 2MOPP for 500Vrms, the devices meet clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7. The risk management process is not applicable to these clauses.
6. Case material: CTI=600V, erosion depth 0.028mm, meets material group I.
7. Evaluated by thermal cycling and other tests for a temperature rating of 125C.
8. The creepage and clearance has been evaluated for altitudes ≤ 2000m, in pollution degree 2, material group III and overvoltage category II except where specified otherwise.

These devices are Component Accepted as components for use in other Certified equipment where the suitability of the combination shall be determined by investigation in the final application.

APPLICABLE REQUIREMENTS

- | | | |
|---------------------------------------|---|--|
| CAN/CSA-C22.2 No 60950-1-07+A1+A2 | - | Information Technology Equipment - Safety - Part 1: General Requirements (Bi-national Standard, with UL 60950-1) |
| CAN/CSA-C22.2 No. 62368-1-14 | - | Audio/video, information and communication technology equipment– Part 1: Safety requirements |
| IEC 60950-1 2 nd Ed.+A1+A2 | - | Information Technology Equipment - Safety - Part 1: General Requirements |



Certificate: 80008667
Project: 80008667

Master Contract: 255632
Date Issued: 2019-11-29

IEC 62368-1:2014 Ed. 2

- Audio/video, information and communication technology equipment– Part 1: Safety requirements

Clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7
of

CAN/CSA C22.2 60601-1:14

- Medical Electrical Equipment Part 1: General requirements for basic safety and general performance (Adopted IEC60601-1:2005 Edition 3.0 +Amendment 1, 2012-07, MOD)

and

IEC 60601-1:2005 Ed 3.0+A1

- Medical Electrical Equipment Part 1: General requirements for basic safety and general performance

ANSI/UL 1577-2015

- Standard for Optical Isolators



Supplement to Certificate of Compliance

Certificate: 80008667

Master Contract: 255632

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80008667	2019-11-29	Original Component Acceptance