



Certificate of Compliance

Certificate:	80218509	Master Contract:	205078
Project:	80218509	Date Issued:	2025-03-19
Issued to:	Analog Devices Inc. 804 Woburn St Wilmington, Massachusetts 01887-3462 United States	Issued by:	Martin Buchanan Martin Buchanan, P. Eng.

Attention: Ivy Mae Tejero

The products listed below are eligible to bear the CSA Mark shown with adjacent indicator ▲



PRODUCTS

Class 9073 30 ELECTRONIC COMPONENTS - Optoisolators and non-optical isolating devices

Model(s)
ADN4620BRIZ, ADN4621BRIZ, ADN4620BRSZ, ADN4621BRSZ

Component Acceptance of Optoisolator Like Devices:



Certificate: 80218509

Master Contract: 205078

Project: 80218509

Date Issued: 2025-03-19

Device	Rating		Standard/Notice and Clauses	Internal		External
	kV	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
(SOIC16 W INCREASED CREEPAGE RI-16-3) ADN4620BRIZ ADN4621BRIZ	7.5	125	CSA 62368-1:19, UPD1:2021 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.8, 5.4.1.5.3, 5.4.9.1, 5.4.1.4 61010-1-12+A1 6.7.1.3, 6.7.2.1 or tbK.1 to K.4, 6.7.2.2.1, 6.7.2.2.2 or tbK.9, A.17, tbK.5x1.6, K.6x1.6, K.7x1.6, 10	-	-	16
(SSOP20 RS-20) ADN4620BRSZ ADN4621BRSZ	3.75	125	60601-1:14 (R2022)+A1+A2 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7 IEC 62368-1:2018 Ed. 3 3 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.8, 5.4.1.5.3, 5.4.9.1, 5.4.1.4 61010-1 3rd Ed+A1 6.7.1.3, 6.7.2.1 or tbK.1 to K.4, 6.7.2.2.1, 6.7.2.2.2 or tbK.9, A.17, tbK.5x1.6, K.6x1.6, K.7x1.6, 10 60601-1 Ed.3+A1+A2 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7 EN 62368-1:2020+A11:2020 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.8, 5.4.1.5.3, 5.4.9.1, 5.4.1.4	-	-	5.3

where Z if used indicates the lead-free version. Suffix EP if present at the end of the base part number indicates lower extended temperature capability to -55C with a NiPdAu lead finish. Further suffix letters or digits differentiate shipping package formats.



Certificate: 80218509

Master Contract: 205078

Project: 80218509

Date Issued: 2025-03-19

Notes:

All Devices

1. Case material: CTI 600V with erosion depth <0.1mm, Material group I.
2. Evaluated by thermal cycling and other tests for a temperature rating of 125C.
3. The creepage and clearance has been evaluated for altitudes $\leq 2000\text{m}$, in pollution degree 2 and overvoltage category II except where specified above. (pollution degree 2, material group III).

(SOIC16 W INCREASED CREEPAGE RI-16-3) RI Devices:

1. The devices meet basic insulation requirements for 1000Vrms and reinforced insulation requirements for 600Vrms for CSA 62368-1:19, UPD1:2021, IEC 62368-1:2018 Ed. 3 and EN 62368-1:2020+A11:2020. (pollution degree 2, material group III)
2. For CSA 61010-1-12+A1 and IEC 61010-1 3rd Ed., the devices meet 1000Vrms for basic insulation and 600V for reinforced insulation based on 61010-1 Cl 14.1 a) for use in 61010-1 end products because they meet the requirements of the 62368-1 evaluation. The risk management process is not applicable to these clauses. (pollution degree 2, material group III)
3. For pollution degree 2, material group I, for CSA 61010-1-12+A1 and IEC 61010-1 3rd Ed., the devices meet 1000Vrms for reinforced insulation based on 61010-1 Cl 14.1 a) for use in 61010-1 end products because they meet the requirements of the 62368-1 evaluation. The risk management process is not applicable to these clauses. (pollution degree 2, material group I)
4. For CSA 60601-1:14 (R2022) and IEC 60601-1:2005 Ed 3.0+A1+A2 for 1 MOPP for 1000Vrms and 2 MOPP for 50Vrms, the devices meet clauses 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7. The risk management process is not applicable to these clauses.

(SSOP20 RS-20) RS Devices:

1. The devices meet basic insulation requirements for 530Vrms and reinforced insulation requirements for 265Vrms for CSA 62368-1:19, UPD1:2021, IEC 62368-1:2018 Ed. 3 and EN62368-1:2020. (pollution degree 2, material group III)
2. For CSA 61010-1-12+A1 and IEC 61010-1 3rd Ed. the devices meet 300Vrms for basic insulation and 150V for reinforced insulation based on 61010-1 Cl 14.1 a) for use in 61010-1 end products because they meet the requirements of the 62368-1 evaluation. The risk management process is not applicable to these clauses. (pollution degree 2, material group III)
3. For CSA 60601-1:14 (R2022) and IEC 60601-1:2005 Ed 3.0+A1+A2 for 1 MOPP for 347Vrms and 2 MOPP for 50Vrms, the devices meet clauses 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7. The risk management process is not applicable to these clauses.

These devices are Component Accepted as components for use in other Certified equipment where the suitability of the combination shall be determined by investigation in the final application.

APPLICABLE REQUIREMENTS

CSA C22.2 No. 62368-1:19 - Third Edition - Including Update No. 1 October 2021 - Audio/video, information and communication technology equipment — Part 1: Safety requirements



Certificate: 80218509

Master Contract: 205078

Project: 80218509

Date Issued: 2025-03-19

CSA C22.2 No. 61010-1-12, UPD1:2015, UPD2:2016, AMD1:2018, UPD3:2023 - Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements - Third Edition; Update No. 1: July 2015; Update No. 2: April 2016; Update No. 3: June 2023

CAN/CSA-C22.2 60601-1:14 (R2022) - Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance (Adopted IEC 60601-1:2005, third edition + Amendment 1:2012)

IEC 62368-1:2018 - Audio/video, information and communication technology equipment – Part 1: Safety requirements - Edition 3.0

IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016 - Amendment 1 - Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements

IEC 60601-1:2005/AMD1:2012/AMD2:2020 - Medical electrical equipment – Part 1: General requirements for basic safety and essential performance - Edition 3.2; Consolidated Reprint; Incorporates Amendment 1: 2012, Corrigenda 1: 12/2012, Corrigenda to Amendment 1: 07/2014, Interpretation 1: 04/2008, Interpretation 2: 01/2009, and Interpretation 3: 05/2013 and Amendment 2: 08/2020

EN 62368-1: 2020/A11:2020 - Audio/video, information and communication technology equipment - Part 1: Safety requirements - Incorporates Amendment A11: 2020



Certificate: 80218509

Project: 80218509

Master Contract: 205078

Date Issued: 2025-03-19

Notes:

Products certified under Class(es) C907330 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca

